



Certificate of Conformity

Certificado de Conformidad / Certificat de Conformité /
Certificado de Conformidade / Certificato di Conformità /
Konformitätszertifikat

Model / Modelo / Modèle / Modello / Gerät:

Serial Number/ Número de Serie / Numéro
de Série / Numero di Serie / Seriennummer:

Test Conditions / Condiciones de la Prueba /
Conditions D'essai / Condições de Teste /
Condizioni di Prova / Test – Bedingungen:

Test Results / Résultat / Resultado /
Risultato / Ergebnis:

Range	Tolerance	
Temperature / Temperatura / Température / Temperatur (°F) / (°C)		
69.8°F / 21°C	±1.0°F / ±0.6°C	✓Pass

Range	Tolerance	
Humidity/ Umidade/ Humidité/ Umidità/ Luftfeuchtigkeit (%)		
33.0%	±2.5%	✓Pass
75.0%	±2.5%	✓Pass
95.0%	±5.0%	✓Pass

Inspector / Inspecteur / Inspetor /
Ispettore / Inspektor

Fieldpiece Instruments, Inc., certifies that this product was calibrated in accordance with applicable calibration procedures and practices using equipment calibrated to National Institute for Standards and Technology (NIST) standards (or equivalent). These procedures are controlled by ISO-9001 and are designed to ensure the instrument meets its published specifications.

Fieldpiece Instruments, Inc. | 1636 West Collins Ave., Orange, CA 92867 USA | +1 (714) 634-1844